



## CERTIFICATE OF ANALYSIS

Collagenase MA, Catalog # 001-2030  
GMP Grade – For *ex vivo* Use Only

**Product:** Aseptically dispensed mixture of purified Class 1 (60%) and Class 2 (40%) Collagenase from *C. histolyticum*

Lot # 122201119	SPECIFICATION	RESULT
Appearance	White lyophilized cake	White lyophilized cake
Specific Collagen Degrading Activity <sup>1</sup>	15,000 – 30,000 U/mg	24,656 Units/mg
Total Collagen Degrading Activity	2.25 – 2.75 M Units/bottle	2,598,742 Units/bottle
Total Protein <sup>2</sup>	Report Only	105.4 mg/bottle
Endotoxin EU/mg protein <sup>3</sup>	< 10.0 EU/mg	1.3 EU/mg

**STORAGE:** -20±5°C

**DATE OF MANUFACTURE:** 11 Nov 2020

**EXPIRY DATE:** Nov 2024

**Signature:**

Andrew Breite, Director of Quality Assurance

4 Jan 2021

<sup>1</sup> McCarthy RC, et. al. Development and Characterization of a Collagen Degradation Assay to Assess Purified Collagenase Used in Islet Isolation. *Transplantation Proceedings* 40 (2008); 339-342

<sup>2</sup> Based on absorbance at 280 nm using an extinction coefficient of  $\epsilon=1.41$

<sup>3</sup> Charles River nexgenEndosafe® PTS Endotoxin Assay