



CERTIFICATE OF ANALYSIS

Cizyme RI, Catalog # 005-1030
For Research Use Only

Product: Aseptically dispensed mixture of purified Class 1 and Class 2 Collagenase from *C. histolyticum* and purified Neutral Protease from *B. polymyxa*

Lot # 245160407	SPECIFICATION	RESULT
Appearance	White lyophilized cake	White lyophilized cake
Total Collagen Degrading Activity ¹	300,000 - 450,000 Units/Vial	391,684 Units
Total Neutral Protease Activity ²	60,000 - 100,000 Units/Vial	79,382 Units
Endotoxin EU/mg protein ³	< 25.0 EU/OD 280nm	2.65 EU/OD 280 nm

STORAGE: < -20°C

DATE OF MANUFACTURE: 7 April 2016
EXPIRY DATE: April 2018

Signature:

Andrew Breite, Director of Quality Assurance

13 April 2016

-
- ¹ McCarthy RC, et. al. Development and Characterization of a Collagen Degradation Assay to Assess Purified Collagenase Used in Islet Isolation. *Transplantation Proceedings* 40 (2008); 339-342
- ² Breite AG, et. al. Tissue Dissociation Enzyme Neutral Protease Assessment. *Transplantation Proceedings* 42 (2010);2052-54
- ³ Charles River Endosafe® PTS Endotoxin Assay

VitaCyte, LLC
1102 Indiana Avenue
Indianapolis, IN 46202-3206
www.vitacyte.com
Phone: 317.917.3457
Fax: 317.917.3459